
Midmark M11 - v1.0.5
Total Cycles: 2826

Name: RITTER2
Sterilizer ID: V2313228
08/31/2023 04:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.7 F 0.0
2:00 159.2 F 0.0
4:00 185.1 F 0.9
6:00 196.8 F 4.6
8:00 215.5 F 8.6
10:00 229.0 F 13.3
12:00 232.6 F 9.5
14:00 237.2 F 9.8
16:00 244.7 F 13.3
18:00 244.3 F 12.6
20:00 256.3 F 19.0
22:00 267.1 F 25.7
24:00 272.5 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.1
1:00 272.3 F 29.2
2:00 272.8 F 29.6
3:00 272.2 F 29.7
4:00 271.7 F 29.0

Min 271.6 F 29.0
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.4 F 0.0
5:00 227.1 F 0.0
10:00 217.8 F 0.0
15:00 181.2 F 0.0
20:00 176.3 F 0.0
25:00 173.2 F 0.0
30:00 170.9 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 24:08
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:02:41
CYCLE COMPLETE

08/31/2023 05:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____