
Midmark M11 - v1.0.5
Total Cycles: 3977

Name: RITTER2
Sterilizer ID: V2313228
09/26/2024 10:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 97.5 F 0.0
2:00 97.7 F 0.0
4:00 118.4 F 0.1
6:00 147.2 F 1.3
8:00 170.9 F 4.3
10:00 193.3 F 7.9
12:00 207.3 F 12.0
14:00 218.3 F 6.4
16:00 226.9 F 11.0
18:00 231.3 F 7.6
20:00 241.4 F 13.5
22:00 242.8 F 12.2
24:00 254.6 F 18.4
26:00 264.7 F 24.9
28:00 271.3 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.7
1:00 272.7 F 29.9
2:00 272.3 F 30.2
3:00 272.2 F 29.7
4:00 272.7 F 30.3

Min 271.8 F 29.5
Max 272.7 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.3 F 0.0
5:00 204.2 F 0.0
10:00 204.5 F 0.0
15:00 190.3 F 0.0
20:00 179.1 F 0.0
25:00 185.7 F 0.0
30:00 198.7 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 28:37
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:07:03
CYCLE COMPLETE

09/26/2024 11:58 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____