
Midmark M11 - v1.0.5
Total Cycles: 1439

Name: RITTER2
Sterilizer ID: V2313228
08/31/2022 12:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.0 F 0.0
2:00 196.2 F 2.7
4:00 204.1 F 7.8
6:00 222.6 F 13.5
8:00 231.4 F 9.7
10:00 234.7 F 8.9
12:00 243.5 F 12.5
14:00 245.0 F 13.2
16:00 257.6 F 20.0
18:00 268.5 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 272.7 F 29.8
2:00 272.1 F 29.1
3:00 271.6 F 29.3
4:00 272.6 F 29.9

Min 271.3 F 28.9
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.9 F 0.0
5:00 235.2 F 0.0
10:00 219.6 F 0.0
15:00 227.9 F 0.0
20:00 223.6 F 0.0
25:00 203.7 F 0.0
30:00 218.7 F 0.0
Drying Complete

FILLING: 1:58
HEATING: 19:15
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:58:20
CYCLE COMPLETE

08/31/2022 01:01 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____