

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3306

Name: RITTER2  
Sterilizer ID: V2313228  
01/17/2024 12:53 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 87.4 F 0.0  
2:00 87.6 F 0.0  
4:00 91.0 F 0.5  
6:00 112.1 F 2.1  
8:00 145.1 F 5.1  
10:00 169.0 F 8.4  
12:00 189.2 F 11.8  
14:00 213.1 F 4.9  
16:00 221.8 F 8.7  
18:00 232.2 F 12.8  
20:00 232.8 F 9.5  
22:00 238.3 F 10.4  
24:00 244.0 F 13.1  
26:00 254.9 F 18.9  
28:00 265.1 F 25.1  
30:00 271.1 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.5  
1:00 272.4 F 30.3  
2:00 272.1 F 29.7  
3:00 271.1 F 29.6  
4:00 272.1 F 30.3

Min 271.1 F 29.4  
Max 272.9 F 30.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.7 F 0.0  
5:00 200.0 F 0.0  
10:00 209.0 F 0.0  
15:00 205.1 F 0.0  
20:00 193.7 F 0.0  
25:00 188.1 F 0.0  
30:00 185.0 F 0.0  
Drying Complete

FILLING: 0:60  
HEATING: 30:02  
STERILIZING: 4:00  
VENTING: 3:34  
DRYING: 30:00  
TOTAL CYCLE: 01:08:35  
CYCLE COMPLETE

01/17/2024 02:01 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_