
Midmark M11 - v1.0.0
Total Cycles: 2579

Name: RITTER 1
Sterilizer ID: V2178198
11/17/2022 04:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.9 F 0.0
2:00 181.7 F 4.6
4:00 199.6 F 9.5
6:00 215.2 F 6.3
8:00 229.6 F 9.3
10:00 235.1 F 9.4
12:00 242.0 F 12.5
14:00 241.6 F 11.3
16:00 255.2 F 18.2
18:00 267.0 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.4 F 29.0
2:00 271.5 F 28.5
3:00 272.4 F 29.1
4:00 271.5 F 28.4

Min 271.4 F 28.3
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.6 F 0.0
5:00 233.2 F 0.0
10:00 232.2 F 0.0
15:00 231.5 F 0.0
20:00 227.2 F 0.0
25:00 229.2 F 0.0
30:00 228.4 F 0.0

Drying Complete

FILLING: 1:28
HEATING: 19:24
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:06
CYCLE COMPLETE

11/17/2022 05:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____