
Midmark M11 - v1.0.0
Total Cycles: 2124

Name: RITTER 1
Sterilizer ID: V2178198
07/13/2022 04:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.8 F 0.0
2:00 176.3 F 2.9
4:00 188.9 F 7.4
6:00 209.6 F 12.5
8:00 223.3 F 7.2
10:00 239.2 F 12.5
12:00 236.1 F 9.4
14:00 237.6 F 9.0
16:00 248.1 F 14.1
18:00 260.2 F 20.9
20:00 271.3 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.4 F 29.3
2:00 271.9 F 28.6
3:00 272.6 F 29.3
4:00 271.8 F 28.8

Min 270.8 F 28.0
Max 273.0 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.3 F 0.0
5:00 245.2 F 0.0
10:00 229.0 F 0.0
15:00 232.4 F 0.0
20:00 229.6 F 0.0
25:00 227.7 F 0.0
30:00 229.3 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 20:33
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:40
CYCLE COMPLETE

07/13/2022 05:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____