

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2265

Name: RITTER2  
Sterilizer ID: V2313228  
04/19/2023 08:42 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 150.9 F 0.0  
2:00 148.5 F 2.0  
4:00 155.1 F 4.7  
6:00 185.2 F 9.0  
8:00 209.4 F 7.5  
10:00 222.4 F 8.8  
12:00 235.5 F 10.8  
14:00 238.9 F 11.2  
16:00 237.0 F 9.7  
18:00 250.9 F 16.4  
20:00 263.0 F 23.3

STERILIZING  
mm:ss Degrees PSI  
0:00 272.1 F 29.2  
1:00 272.2 F 29.4  
2:00 272.5 F 29.6  
3:00 272.0 F 29.8  
4:00 271.2 F 29.2

Min 271.1 F 29.1  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 172.3 F 0.0  
5:00 237.1 F 0.0  
10:00 227.4 F 0.0  
15:00 197.2 F 0.0  
20:00 189.9 F 0.0  
25:00 189.0 F 0.0  
30:00 188.6 F 0.0  
Drying Complete

FILLING: 0:60  
HEATING: 21:58  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:00:09  
CYCLE COMPLETE

04/19/2023 09:42 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_