
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000266
12:10 22/07/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:05
PRESSURIZING 3:08
STERILIZING 4:11
136.4 C 220kPa 4:11
135.9 C 217kPa 4:41
136.4 C 222kPa 5:11
136.4 C 224kPa 5:41
136.4 C 225kPa 6:11
136.4 C 224kPa 6:41
136.2 C 222kPa 7:11
136.2 C 222kPa 7:41
136.2 C 222kPa 7:41
VENTING 7:42
AIR DRYING 8:33
CYCLE COMPLETE 8:37

Digital Signature #
A46E36AA1E619268
