

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2309

Name: RITTER 1  
Sterilizer ID: V2178198  
09/12/2022 10:25 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 168.4 F 0.0  
2:00 169.2 F 3.5  
4:00 188.1 F 7.5  
6:00 205.0 F 12.0  
8:00 217.4 F 5.7  
10:00 229.3 F 10.7  
12:00 230.2 F 7.1  
14:00 242.6 F 12.8  
16:00 240.5 F 10.9  
18:00 253.3 F 17.3  
20:00 264.5 F 23.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.6  
1:00 272.6 F 29.4  
2:00 271.5 F 28.5  
3:00 272.5 F 29.5  
4:00 272.0 F 28.9

Min 271.4 F 28.5  
Max 272.6 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.9 F 0.0  
5:00 229.0 F 0.0  
10:00 236.9 F 0.0  
15:00 236.0 F 0.0  
20:00 227.2 F 0.0  
25:00 227.3 F 0.0  
30:00 215.3 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 21:60  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:00:04  
CYCLE COMPLETE

09/12/2022 11:25 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_