

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1014

Name: RITTER2  
Sterilizer ID: V2313228  
04/20/2022 12:59 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 202.9 F 0.0  
2:00 196.0 F 2.8  
4:00 214.0 F 8.8  
6:00 230.4 F 9.1  
8:00 239.7 F 11.6  
10:00 238.9 F 10.3  
12:00 238.2 F 9.7  
14:00 253.0 F 16.9  
16:00 265.1 F 24.2

STERILIZING  
mm:ss Degrees PSI  
0:00 272.1 F 28.5  
1:00 272.6 F 29.4  
2:00 272.2 F 28.6  
3:00 272.0 F 29.1  
4:00 272.7 F 29.1  
  
Min 271.3 F 28.4  
Max 272.9 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.2 F 0.0  
5:00 229.7 F 0.0  
10:00 223.2 F 0.0  
15:00 226.1 F 0.0  
20:00 226.4 F 0.0  
25:00 219.5 F 0.0  
30:00 190.0 F 0.0  
Drying Complete

FILLING: 2:51  
HEATING: 17:39  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:57:42  
CYCLE COMPLETE

04/20/2022 01:56 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_