

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3684

Name: RITTER 1  
Sterilizer ID: V2178198  
08/25/2023 12:57 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 100.8 F 0.0  
2:00 101.2 F 0.2  
4:00 124.6 F 2.2  
6:00 149.0 F 4.7  
8:00 168.5 F 7.9  
10:00 191.2 F 11.7  
12:00 214.4 F 4.0  
14:00 218.5 F 7.8  
16:00 229.0 F 11.8  
18:00 230.6 F 7.3  
20:00 241.5 F 12.4  
22:00 238.6 F 10.0  
24:00 250.5 F 15.8  
26:00 261.1 F 21.7  
28:00 270.8 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.6  
1:00 272.4 F 29.3  
2:00 272.1 F 28.7  
3:00 272.1 F 29.2  
4:00 272.7 F 29.0

Min 271.2 F 28.5  
Max 273.0 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.2 F 0.0  
5:00 227.7 F 0.0  
10:00 213.7 F 0.0  
15:00 218.9 F 0.0  
20:00 192.1 F 0.0  
25:00 193.2 F 0.0  
30:00 184.6 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 28:40  
STERILIZING: 4:00  
VENTING: 3:28  
DRYING: 30:00  
TOTAL CYCLE: 01:06:59  
CYCLE COMPLETE

08/25/2023 02:03 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_