
Midmark M11 - v1.0.5
Total Cycles: 1256

Name: RITTER2
Sterilizer ID: V2313228
07/05/2022 11:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.2 F 0.0
2:00 174.2 F 3.5
4:00 191.1 F 8.2
6:00 222.5 F 10.6
8:00 230.1 F 9.3
10:00 234.4 F 9.0
12:00 242.4 F 12.8
14:00 241.9 F 11.8
16:00 254.7 F 18.4
18:00 265.4 F 25.0
20:00 271.2 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 272.4 F 29.5
1:00 272.6 F 29.9
2:00 272.0 F 30.1
3:00 271.7 F 29.5
4:00 272.5 F 29.8

Min 271.5 F 29.4
Max 272.6 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.4 F 0.0
5:00 230.2 F 0.0
10:00 218.0 F 0.0
15:00 220.9 F 0.0
20:00 219.4 F 0.0
25:00 222.8 F 0.0
30:00 220.4 F 0.0
Drying Complete

FILLING: 2:13
HEATING: 20:26
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:53
CYCLE COMPLETE

07/05/2022 12:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____