
Midmark M11 - v1.0.5
Total Cycles: 1313

Name: RITTER2
Sterilizer ID: V2313228
07/18/2022 01:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 122.2 F 0.0
2:00 126.1 F 0.0
4:00 148.8 F 0.3
6:00 185.2 F 1.4
8:00 206.6 F 5.6
10:00 218.1 F 9.8
12:00 230.7 F 9.0
14:00 236.0 F 10.5
16:00 233.6 F 8.2
18:00 242.2 F 11.5
20:00 246.0 F 13.6
22:00 258.3 F 20.3
24:00 269.4 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.5 F 29.7
2:00 271.7 F 28.9
3:00 272.0 F 29.6
4:00 272.3 F 29.4

Min 271.3 F 28.8
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.5 F 0.0
5:00 235.8 F 0.0
10:00 236.6 F 0.0
15:00 229.7 F 0.0
20:00 226.6 F 0.0
25:00 227.4 F 0.0
30:00 228.5 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 25:04
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:03:13
CYCLE COMPLETE

07/18/2022 02:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____