
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 002517
13:43 25/10/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:39
PRESSURIZING 3:40
STERILIZING 4:37
136.0 C 220kPa 4:37
136.3 C 224kPa 5:07
136.4 C 224kPa 5:37
136.1 C 222kPa 6:07
136.4 C 224kPa 6:37
136.4 C 225kPa 7:07
136.1 C 223kPa 7:37
136.3 C 223kPa 8:07
136.3 C 223kPa 8:07
VENTING 8:07
AIR DRYING 9:03
CYCLE COMPLETE 9:04

Digital Signature #
378DE2F026B5B7EB
