

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1623

Name: RITTER2  
Sterilizer ID: V2313228  
10/19/2022 10:57 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 187.2 F 0.0  
2:00 186.1 F 3.7  
4:00 190.4 F 8.2  
6:00 210.1 F 12.2  
8:00 224.8 F 8.8  
10:00 236.5 F 10.8  
12:00 240.6 F 12.5  
14:00 240.7 F 11.5  
16:00 254.0 F 18.3  
18:00 265.5 F 25.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.4  
1:00 272.5 F 30.0  
2:00 271.7 F 29.4  
3:00 271.8 F 30.0  
4:00 272.4 F 29.9  
  
Min 271.2 F 29.2  
Max 272.8 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.7 F 0.0  
5:00 224.4 F 0.0  
10:00 238.8 F 0.0  
15:00 245.4 F 0.0  
20:00 230.9 F 0.0  
25:00 222.0 F 0.0  
30:00 220.0 F 0.0  
Drying Complete

FILLING: 1:43  
HEATING: 19:45  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:58:42  
CYCLE COMPLETE

10/19/2022 11:55 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_