
Midmark M11 - v1.0.5
Total Cycles: 3223

Name: RITTER2
Sterilizer ID: V2313228
12/15/2023 11:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.1 F 0.0
2:00 159.1 F 1.6
4:00 175.3 F 5.5
6:00 196.4 F 10.2
8:00 222.9 F 5.7
10:00 229.7 F 10.6
12:00 231.1 F 7.5
14:00 242.1 F 11.9
16:00 245.3 F 13.1
18:00 257.8 F 19.8
20:00 268.9 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.0
1:00 272.6 F 29.5
2:00 271.8 F 29.4
3:00 273.1 F 29.9
4:00 272.2 F 30.1

Min 270.9 F 28.8
Max 273.6 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 160.7 F 0.0
5:00 208.0 F 0.0
10:00 201.9 F 0.0
15:00 189.6 F 0.0
20:00 187.4 F 0.0
25:00 185.2 F 0.0
30:00 182.5 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 21:09
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 00:59:47
CYCLE COMPLETE

12/15/2023 12:42 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____