

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1176

Name: RITTER 1  
Sterilizer ID: V2178198  
09/15/2021 10:27 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 145.7 F 0.0  
2:00 150.1 F 1.6  
4:00 175.0 F 4.6  
6:00 195.5 F 8.5  
8:00 214.1 F 13.2  
10:00 222.5 F 7.4  
12:00 235.2 F 12.5  
14:00 236.4 F 9.8  
16:00 237.4 F 8.9  
18:00 249.0 F 14.9  
20:00 261.0 F 21.6  
22:00 271.8 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.5  
1:00 272.8 F 29.3  
2:00 272.0 F 28.6  
3:00 272.0 F 28.9  
4:00 272.8 F 29.2

Min 271.5 F 28.4  
Max 273.0 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.8 F 0.0  
5:00 208.2 F 0.0  
10:00 210.6 F 0.0  
15:00 205.3 F 0.0  
20:00 209.3 F 0.0  
25:00 207.3 F 0.0  
30:00 222.8 F 0.0  
Drying Complete

FILLING: 1:14  
HEATING: 22:36  
STERILIZING: 4:00  
VENTING: 3:21  
DRYING: 30:00  
TOTAL CYCLE: 01:01:10  
CYCLE COMPLETE

09/15/2021 11:28 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01