
Midmark M11 - v1.0.0
Total Cycles: 2162

Name: RITTER 1
Sterilizer ID: V2178198
07/21/2022 11:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.9 F 0.0
2:00 179.4 F 3.5
4:00 190.3 F 8.0
6:00 208.7 F 13.0
8:00 225.4 F 6.9
10:00 236.1 F 12.1
12:00 235.5 F 9.2
14:00 238.5 F 9.6
16:00 247.0 F 13.9
18:00 259.4 F 20.6
20:00 270.1 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.5 F 29.2
2:00 271.5 F 28.5
3:00 272.3 F 29.3
4:00 272.4 F 28.9

Min 271.2 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.7 F 0.0
5:00 230.9 F 0.0
10:00 225.4 F 0.0
15:00 218.4 F 0.0
20:00 222.7 F 0.0
25:00 232.1 F 0.0
30:00 204.6 F 0.0
Drying Complete

FILLING: 1:31
HEATING: 20:55
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:41
CYCLE COMPLETE

07/21/2022 12:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____