

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4464

Name: RITTER2  
Sterilizer ID: V2313228  
03/27/2025 11:19 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 128.0 F 0.0  
2:00 130.0 F 1.0  
4:00 132.0 F 3.3  
6:00 147.7 F 6.5  
8:00 168.1 F 10.5  
10:00 214.8 F 5.0  
12:00 216.5 F 7.1  
14:00 229.8 F 11.3  
16:00 230.4 F 7.8  
18:00 241.4 F 13.2  
20:00 241.1 F 11.7  
22:00 252.4 F 17.4  
24:00 262.3 F 23.5  
26:00 271.8 F 29.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.8  
1:00 271.8 F 30.5  
2:00 272.7 F 30.5  
3:00 272.9 F 30.6  
4:00 272.2 F 30.9

Min 271.0 F 29.6  
Max 273.4 F 30.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.4 F 0.0  
5:00 185.8 F 0.0  
10:00 194.7 F 0.0  
15:00 192.5 F 0.0  
20:00 190.7 F 0.0  
25:00 189.3 F 0.0  
30:00 187.8 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 26:42  
STERILIZING: 4:00  
VENTING: 3:55  
DRYING: 30:00  
TOTAL CYCLE: 01:05:31  
CYCLE COMPLETE

03/27/2025 12:24 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_