
Midmark M11 - v1.0.0
Total Cycles: 4498

Name: RITTER 1
Sterilizer ID: V2178198
04/26/2024 02:56 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.8 F 0.0
2:00 154.3 F 2.7
4:00 173.2 F 6.1
6:00 191.6 F 10.0
8:00 222.3 F 7.3
10:00 222.7 F 8.4
12:00 235.0 F 13.1
14:00 235.5 F 9.6
16:00 238.4 F 9.4
18:00 246.7 F 13.4
20:00 258.1 F 19.5
22:00 268.1 F 25.9
24:00 273.5 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 273.1 F 29.4
2:00 272.1 F 28.8
3:00 271.2 F 28.7
4:00 272.4 F 29.4

Min 271.1 F 28.5
Max 273.4 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.5 F 0.0
5:00 242.1 F 0.0
10:00 192.3 F 0.0
15:00 189.1 F 0.0
20:00 187.3 F 0.0
25:00 185.7 F 0.0
30:00 183.8 F 0.0
Drying Complete

FILLING: 1:20
HEATING: 24:21
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:03:20
CYCLE COMPLETE

04/26/2024 03:59 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____