
Midmark M11 - v1.0.0
Total Cycles: 3416

Name: RITTER 1
Sterilizer ID: V2178198
06/20/2023 02:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.9 F 0.0
2:00 169.7 F 1.8
4:00 188.4 F 5.6
6:00 214.6 F 10.5
8:00 221.3 F 5.2
10:00 234.7 F 10.8
12:00 233.1 F 8.0
14:00 241.4 F 10.8
16:00 246.5 F 13.4
18:00 259.2 F 20.3
20:00 270.5 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.6 F 28.9
2:00 271.4 F 28.1
3:00 272.8 F 29.1
4:00 272.0 F 28.4

Min 271.4 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.3 F 0.0
5:00 208.9 F 0.0
10:00 222.3 F 0.0
15:00 215.4 F 0.0
20:00 216.1 F 0.0
25:00 218.2 F 0.0
30:00 220.3 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 20:49
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:25
CYCLE COMPLETE

06/20/2023 03:24 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____