
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000914
15:52 07/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:38
PRESSURIZING 2:45
STERILIZING 4:04
136.4 C 233kPa 4:04
136.2 C 231kPa 4:34
136.4 C 232kPa 5:04
136.4 C 230kPa 5:34
136.4 C 229kPa 6:04
136.4 C 232kPa 6:34
136.4 C 233kPa 7:04
136.4 C 233kPa 7:34
136.4 C 233kPa 7:34
VENTING 7:34
AIR DRYING 8:27
CYCLE COMPLETE 23:27

Digital Signature #
1E164581582025C1
