
Midmark M11 - v1.0.5
Total Cycles: 3080

Name: RITTER2
Sterilizer ID: V2313228
11/03/2023 08:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.3 F 0.0
2:00 73.0 F 0.0
4:00 98.6 F 0.4
6:00 124.6 F 1.9
8:00 148.3 F 5.1
10:00 173.1 F 8.8
12:00 192.1 F 12.8
14:00 208.6 F 4.6
16:00 220.3 F 8.2
18:00 231.7 F 12.4
20:00 232.4 F 8.6
22:00 241.9 F 11.8
24:00 243.9 F 12.6
26:00 256.2 F 19.0
28:00 267.0 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 272.7 F 29.4
2:00 272.2 F 29.5
3:00 272.4 F 29.1
4:00 272.5 F 29.7

Min 271.6 F 28.8
Max 273.0 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.6 F 0.0
5:00 192.0 F 0.0
10:00 191.2 F 0.0
15:00 190.8 F 0.0
20:00 191.7 F 0.0
25:00 191.8 F 0.0
30:00 193.4 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 29:28
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:07:29
CYCLE COMPLETE

11/03/2023 10:03 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____