
Midmark M11 - v1.0.0
Total Cycles: 4215

Name: RITTER 1
Sterilizer ID: V2178198
01/19/2024 11:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.0 F 0.0
2:00 166.9 F 2.6
4:00 178.6 F 6.2
6:00 205.6 F 10.9
8:00 220.3 F 4.4
10:00 227.3 F 9.3
12:00 237.5 F 10.4
14:00 241.7 F 11.2
16:00 238.2 F 9.4
18:00 251.4 F 15.7
20:00 262.4 F 22.0
22:00 272.7 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.6 F 28.9
2:00 272.5 F 28.8
3:00 272.4 F 28.9
4:00 272.4 F 29.0

Min 270.7 F 27.9
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.8 F 0.0
5:00 229.7 F 0.0
10:00 225.0 F 0.0
15:00 202.7 F 0.0
20:00 226.6 F 0.0
25:00 230.8 F 0.0
30:00 232.5 F 0.0
Drying Complete

FILLING: 1:31
HEATING: 22:17
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:01:32
CYCLE COMPLETE

01/19/2024 12:33 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____