
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001705
10:35 21/02/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:30
PRESSURIZING 2:34
STERILIZING 3:39
136.0 C 220kPa 3:39
136.4 C 225kPa 4:09
136.3 C 223kPa 4:39
136.4 C 224kPa 5:09
136.2 C 223kPa 5:39
136.2 C 222kPa 6:09
136.4 C 224kPa 6:39
136.4 C 224kPa 7:09
136.4 C 224kPa 7:09
VENTING 7:09
AIR DRYING 8:00
CYCLE COMPLETE 17:13

Digital Signature #
05D6630BE294056E
