
Midmark M11 - v1.0.0
Total Cycles: 5240

Name: RITTER 1
Sterilizer ID: V2178198
01/13/2025 02:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.6 F 0.0
2:00 150.1 F 0.1
4:00 167.7 F 1.3
6:00 193.8 F 5.4
8:00 208.3 F 10.0
10:00 225.5 F 7.3
12:00 229.9 F 9.1
14:00 239.3 F 10.7
16:00 241.3 F 11.7
18:00 239.0 F 10.2
20:00 252.8 F 16.9
22:00 264.3 F 23.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.1 F 28.7
2:00 272.3 F 29.1
3:00 271.5 F 28.3
4:00 272.5 F 29.0

Min 271.4 F 28.3
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.7 F 0.0
5:00 206.9 F 0.0
10:00 209.5 F 0.0
15:00 217.2 F 0.0
20:00 232.2 F 0.0
25:00 230.3 F 0.0
30:00 228.5 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 23:55
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:02:17
CYCLE COMPLETE

01/13/2025 03:12 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____