

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001784  
9:30 12/06/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:00  
PRESSURIZING 3:06  
STERILIZING 4:18  
136.2 C 220kPa 4:18  
136.2 C 223kPa 4:48  
136.4 C 224kPa 5:18  
136.2 C 222kPa 5:48  
136.4 C 224kPa 6:18  
136.4 C 222kPa 6:48  
136.1 C 222kPa 7:18  
136.2 C 220kPa 7:48  
136.2 C 220kPa 7:48  
VENTING 7:48  
AIR DRYING 8:38  
CYCLE COMPLETE 38:38

-----  
Digital Signature #  
5B41402D6907554C  
-----