
Midmark M11 - v1.0.0
Total Cycles: 4754

Name: RITTER 1
Sterilizer ID: V2178198
07/22/2024 11:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.0 F 0.0
2:00 163.3 F 2.1
4:00 166.7 F 5.7
6:00 185.3 F 9.4
8:00 202.7 F 13.5
10:00 214.3 F 5.9
12:00 226.4 F 10.1
14:00 232.8 F 8.8
16:00 238.6 F 10.9
18:00 235.3 F 8.5
20:00 248.7 F 14.7
22:00 259.7 F 20.8
24:00 269.8 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.5 F 29.1
2:00 271.6 F 28.4
3:00 271.7 F 28.9
4:00 272.3 F 29.0

Min 271.2 F 28.3
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.6 F 0.0
5:00 218.2 F 0.0
10:00 210.0 F 0.0
15:00 212.4 F 0.0
20:00 217.2 F 0.0
25:00 230.8 F 0.0
30:00 231.1 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 24:59
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:03:42
CYCLE COMPLETE

07/22/2024 12:39 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____