
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000341
15:11 18/08/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:32
PRESSURIZING 3:35
STERILIZING 4:40
136.0 C 220kPa 4:40
136.4 C 224kPa 5:10
136.4 C 223kPa 5:40
136.2 C 223kPa 6:10
136.2 C 223kPa 6:40
136.2 C 221kPa 7:10
136.2 C 223kPa 7:40
136.3 C 223kPa 8:10
136.3 C 223kPa 8:10
VENTING 8:10
AIR DRYING 9:03
CYCLE COMPLETE 9:11

Digital Signature #
ED753E09F506E86D
