
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002935
15:07 11/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:49
PRESSURIZING 2:50
STERILIZING 3:38
136.0 C 220kPa 3:38
136.0 C 223kPa 4:08
135.9 C 218kPa 4:38
136.1 C 221kPa 5:08
136.4 C 225kPa 5:38
136.4 C 222kPa 6:08
136.0 C 217kPa 6:38
136.3 C 224kPa 7:08
136.3 C 224kPa 7:08
VENTING 7:08
AIR DRYING 8:09
CYCLE COMPLETE 8:11

Digital Signature #
6E68C7E57ABC236B
