

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2514

Name: RITTER2  
Sterilizer ID: V2313228  
06/14/2023 10:45 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 196.5 F 0.0  
2:00 193.1 F 3.6  
4:00 219.9 F 8.8  
6:00 224.9 F 8.6  
8:00 234.5 F 10.4  
10:00 232.5 F 7.9  
12:00 240.0 F 10.9  
14:00 246.3 F 13.9  
16:00 258.9 F 20.8  
18:00 270.2 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 272.1 F 29.0  
1:00 272.2 F 29.8  
2:00 272.1 F 29.2  
3:00 271.7 F 29.5  
4:00 272.4 F 29.6  
  
Min 271.4 F 28.9  
Max 272.6 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.0 F 0.0  
5:00 221.2 F 0.0  
10:00 237.0 F 0.0  
15:00 232.2 F 0.0  
20:00 227.0 F 0.0  
25:00 227.4 F 0.0  
30:00 227.9 F 0.0  
Drying Complete

FILLING: 1:59  
HEATING: 18:44  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:57:59  
CYCLE COMPLETE

06/14/2023 11:42 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_