
Midmark M11 - v1.0.0
Total Cycles: 1993

Name: RITTER 1
Sterilizer ID: V2178198
06/06/2022 10:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 129.2 F 0.0
2:00 132.1 F 1.8
4:00 144.7 F 4.7
6:00 178.7 F 8.6
8:00 199.8 F 13.1
10:00 217.4 F 6.7
12:00 230.8 F 11.4
14:00 229.3 F 6.9
16:00 240.9 F 12.3
18:00 238.3 F 9.8
20:00 250.8 F 15.7
22:00 261.6 F 22.0
24:00 271.7 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.3 F 29.1
2:00 271.3 F 28.6
3:00 272.3 F 29.1
4:00 271.2 F 28.5

Min 271.2 F 28.4
Max 272.5 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.6 F 0.0
5:00 209.7 F 0.0
10:00 209.9 F 0.0
15:00 235.6 F 0.0
20:00 235.3 F 0.0
25:00 228.2 F 0.0
30:00 223.4 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 24:36
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:02:35
CYCLE COMPLETE

06/06/2022 11:47 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____