
Surrey, BC
Canada

STATIM 5000 S501R711
SN 511018F00076
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000059
16:14 19/01/2021
WRAPPED
134 C for 6 min.
CYCLE START 0:00
CONDITIONING 5:40
CYCLE ABORTED 6:18
STOP BUTTON PRESSED
NOT STERILE
END TIME 6:18

Digital Signature #
7C1BFB2E2B44DFEC
