

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5448

Name: RITTER 1  
Sterilizer ID: V2178198  
03/26/2025 11:45 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 151.5 F 0.0  
2:00 155.6 F 2.5  
4:00 167.8 F 5.8  
6:00 191.3 F 9.6  
8:00 208.5 F 8.1  
10:00 219.3 F 6.4  
12:00 230.4 F 10.7  
14:00 230.9 F 7.4  
16:00 241.8 F 12.1  
18:00 239.7 F 10.3  
20:00 251.9 F 16.4  
22:00 262.8 F 22.8  
24:00 272.7 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.5  
1:00 272.7 F 29.3  
2:00 272.3 F 28.9  
3:00 271.5 F 28.4  
4:00 271.9 F 29.0

Min 271.4 F 28.4  
Max 272.9 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.8 F 0.0  
5:00 199.0 F 0.0  
10:00 196.5 F 0.0  
15:00 191.6 F 0.0  
20:00 194.9 F 0.0  
25:00 199.6 F 0.0  
30:00 218.5 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 24:23  
STERILIZING: 4:00  
VENTING: 3:59  
DRYING: 30:00  
TOTAL CYCLE: 01:03:27  
CYCLE COMPLETE

03/26/2025 12:48 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_