
Midmark M11 - v1.0.5
Total Cycles: 3332

Name: RITTER2
Sterilizer ID: V2313228
01/29/2024 02:53 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 120.0 F 0.0
2:00 123.7 F 0.9
4:00 150.0 F 4.2
6:00 180.5 F 8.9
8:00 201.3 F 7.8
10:00 221.8 F 7.3
12:00 238.5 F 12.4
14:00 236.8 F 9.7
16:00 237.2 F 9.0
18:00 249.0 F 15.0
20:00 262.0 F 22.3
22:00 272.3 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 272.4 F 28.9
2:00 272.7 F 29.5
3:00 271.8 F 28.6
4:00 272.7 F 29.2

Min 271.7 F 28.5
Max 272.9 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.2 F 0.0
5:00 216.4 F 0.0
10:00 193.4 F 0.0
15:00 198.1 F 0.0
20:00 203.8 F 0.0
25:00 208.4 F 0.0
30:00 207.4 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 22:11
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:00:21
CYCLE COMPLETE

01/29/2024 03:53 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____