
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 007932
10:56 07/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:09
PRESSURIZING 3:10
STERILIZING 4:16
136.5 C 222kPa 4:16
136.0 C 223kPa 4:46
136.2 C 222kPa 5:16
136.4 C 223kPa 5:46
136.4 C 225kPa 6:16
136.1 C 222kPa 6:46
136.2 C 218kPa 7:16
136.4 C 223kPa 7:46
136.4 C 223kPa 7:46
VENTING 7:46
AIR DRYING 8:44
CYCLE COMPLETE 29:11

Digital Signature #
54BD443F4A8291E9
