
Midmark M11 - v1.0.5
Total Cycles: 251

Name: RITTER2
Sterilizer ID: V2313228
07/27/2021 09:41 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 124.8 F 0.0
2:00 125.2 F 1.2
4:00 145.1 F 3.9
6:00 171.9 F 7.7
8:00 193.6 F 12.3
10:00 217.1 F 6.3
12:00 232.0 F 10.9
14:00 231.9 F 7.2
16:00 243.2 F 13.0
18:00 241.7 F 11.6
20:00 254.1 F 17.7
22:00 265.0 F 24.4
24:00 271.3 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.6
1:00 272.2 F 29.9
2:00 272.1 F 30.1
3:00 272.5 F 30.0
4:00 272.3 F 30.0

Min 271.5 F 29.5
Max 273.1 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 166.4 F 0.0
5:00 237.0 F 0.0
10:00 214.7 F 0.0
15:00 216.0 F 0.0
20:00 164.7 F 0.0
25:00 159.8 F 0.0
30:00 157.1 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 24:28
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:02:49
CYCLE COMPLETE

07/27/2021 10:43 AM
ACCEPTED: X
REJECTED:
Operator: U01