
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001412
12:52 17/03/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:12
PRESSURIZING 3:20
STERILIZING 4:29
136.0 C 220kPa 4:29
136.4 C 225kPa 4:59
136.3 C 225kPa 5:29
136.3 C 224kPa 5:59
136.0 C 222kPa 6:29
136.2 C 223kPa 6:59
136.3 C 224kPa 7:29
136.4 C 224kPa 7:59
136.4 C 224kPa 7:59
VENTING 7:59
AIR DRYING 8:51
CYCLE COMPLETE 23:51

Digital Signature #
F07F7CEDBF787C29
