
Midmark M11 - v1.0.0
Total Cycles: 2842

Name: RITTER 1
Sterilizer ID: V2178198
02/01/2023 01:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 120.6 F 0.0
2:00 131.1 F 2.1
4:00 168.7 F 5.2
6:00 186.9 F 9.3
8:00 210.7 F 9.7
10:00 217.5 F 6.6
12:00 227.7 F 11.3
14:00 230.6 F 7.2
16:00 241.5 F 12.9
18:00 241.0 F 11.0
20:00 253.9 F 17.4
22:00 265.3 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.5 F 29.1
2:00 271.6 F 28.8
3:00 272.4 F 29.2
4:00 271.5 F 28.5

Min 271.5 F 28.5
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.3 F 0.0
5:00 217.1 F 0.0
10:00 224.6 F 0.0
15:00 228.9 F 0.0
20:00 235.1 F 0.0
25:00 225.9 F 0.0
30:00 224.2 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 23:51
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:02:07
CYCLE COMPLETE

02/01/2023 02:39 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____