
Midmark M11 - v1.0.5
Total Cycles: 2112

Name: RITTER2
Sterilizer ID: V2313228
03/02/2023 08:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	80.8 F	0.0
2:00	84.0 F	0.0
4:00	127.1 F	0.0
6:00	157.9 F	0.4
8:00	188.8 F	1.9
10:00	209.1 F	5.3
12:00	223.1 F	9.1
14:00	231.9 F	13.1
16:00	230.1 F	8.3
18:00	240.4 F	12.9
20:00	237.0 F	10.2
22:00	237.0 F	8.8

C103: HEATUP MODE
STOP PRESSED
CHAMBER TEMP 246F
CHAMBER PRESS 13.6PSI
** ITEMS NOT STERILE **

03/02/2023 09:01 AM