
Midmark M11 - v1.0.0
Total Cycles: 4290

Name: RITTER 1
Sterilizer ID: V2178198
02/09/2024 02:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 110.7 F 0.0
2:00 132.4 F 1.2
4:00 132.6 F 4.1
6:00 163.1 F 7.7
8:00 186.0 F 11.9
10:00 209.7 F 4.3
12:00 215.7 F 9.0
14:00 235.3 F 11.2
16:00 236.5 F 10.0
18:00 235.6 F 8.5
20:00 246.9 F 13.9
22:00 258.8 F 20.3
24:00 269.8 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.5
1:00 272.6 F 29.1
2:00 271.1 F 28.5
3:00 272.5 F 29.2
4:00 271.5 F 28.5

Min 271.1 F 28.4
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.6 F 0.0
5:00 231.6 F 0.0
10:00 229.8 F 0.0
15:00 226.1 F 0.0
20:00 228.4 F 0.0
25:00 230.0 F 0.0
30:00 226.9 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 24:59
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:03:15
CYCLE COMPLETE

02/09/2024 03:35 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____