
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 006592
9:14 13/11/2023

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:33
PRESSURIZING 2:35
STERILIZING 3:43
136.4 C 220kPa 3:43
136.4 C 219kPa 4:13
136.4 C 223kPa 4:43
136.4 C 222kPa 5:13
136.4 C 221kPa 5:43
136.3 C 221kPa 6:13
136.4 C 223kPa 6:43
136.3 C 221kPa 7:13
136.3 C 221kPa 7:13
VENTING 7:13
AIR DRYING 8:02
CYCLE COMPLETE 8:12

Digital Signature #
87A3454ADB4AD1F5
