

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3839

Name: RITTER2  
Sterilizer ID: V2313228  
08/08/2024 09:57 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 167.5 F 0.0  
2:00 163.3 F 1.6  
4:00 168.0 F 4.7  
6:00 192.9 F 8.9  
8:00 212.9 F 12.9  
10:00 219.9 F 6.6  
12:00 230.9 F 10.9  
14:00 232.1 F 7.5  
16:00 242.6 F 12.9  
18:00 242.1 F 11.7  
20:00 253.8 F 17.6  
22:00 264.0 F 23.9  
24:00 271.5 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.9  
1:00 272.1 F 28.9  
2:00 272.2 F 29.3  
3:00 271.9 F 29.5  
4:00 271.2 F 28.8

Min 271.1 F 28.8  
Max 272.6 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 169.3 F 0.0  
5:00 233.1 F 0.0  
10:00 214.3 F 0.0  
15:00 186.7 F 0.0  
20:00 175.8 F 0.0  
25:00 172.4 F 0.0  
30:00 172.2 F 0.0  
Drying Complete

FILLING: 1:30  
HEATING: 24:02  
STERILIZING: 4:00  
VENTING: 3:40  
DRYING: 30:00  
TOTAL CYCLE: 01:03:12  
CYCLE COMPLETE

08/08/2024 11:00 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_