
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002729
12:18 24/11/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:29
PRESSURIZING 2:33
STERILIZING 3:28
136.4 C 222kPa 3:28
136.0 C 222kPa 3:58
136.4 C 222kPa 4:28
136.0 C 222kPa 4:58
136.2 C 224kPa 5:28
136.4 C 225kPa 5:58
136.4 C 224kPa 6:28
136.2 C 223kPa 6:58
136.2 C 223kPa 6:58
VENTING 6:58
AIR DRYING 7:52
CYCLE COMPLETE 7:53

Digital Signature #
620E94E1CBE1643E
