
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001748
16:11 17/04/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:10
PRESSURIZING 2:17
STERILIZING 3:15
136.0 C 220kPa 3:15
136.4 C 226kPa 3:45
136.4 C 226kPa 4:15
136.4 C 226kPa 4:45
136.4 C 226kPa 5:15
136.4 C 226kPa 5:45
136.4 C 227kPa 6:15
136.4 C 227kPa 6:45
136.4 C 227kPa 6:45
VENTING 6:46
AIR DRYING 7:34
CYCLE COMPLETE 8:05

Digital Signature #
5794FF5CE8D44E0A
