

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1652

Name: RITTER2  
Sterilizer ID: V2313228  
10/25/2022 10:24 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 156.4 F 0.0  
2:00 161.5 F 2.8  
4:00 181.4 F 7.3  
6:00 200.9 F 12.8  
8:00 218.6 F 7.6  
10:00 235.1 F 13.2  
12:00 236.9 F 10.9  
14:00 236.9 F 9.6  
16:00 250.6 F 16.5  
18:00 262.7 F 23.5  
20:00 272.4 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.5  
1:00 272.7 F 30.0  
2:00 271.6 F 29.5  
3:00 272.7 F 30.1  
4:00 271.9 F 29.4

Min 271.6 F 29.3  
Max 272.8 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.7 F 0.0  
5:00 199.2 F 0.0  
10:00 214.5 F 0.0  
15:00 220.8 F 0.0  
20:00 219.7 F 0.0  
25:00 222.6 F 0.0  
30:00 220.5 F 0.0  
Drying Complete

FILLING: 1:57  
HEATING: 20:12  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:59:19  
CYCLE COMPLETE

10/25/2022 11:23 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_