
Midmark M11 - v1.0.0
Total Cycles: 2974

Name: RITTER 1
Sterilizer ID: V2178198
03/06/2023 11:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.8 F 0.0
2:00 186.3 F 3.6
4:00 196.2 F 8.1
6:00 214.7 F 12.2
8:00 227.8 F 8.3
10:00 240.8 F 12.1
12:00 241.2 F 11.4
14:00 239.4 F 9.9
16:00 253.3 F 16.9
18:00 265.2 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.6 F 28.9
2:00 271.5 F 28.2
3:00 272.6 F 29.2
4:00 272.1 F 28.6

Min 271.3 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.2 F 0.0
5:00 239.6 F 0.0
10:00 235.4 F 0.0
15:00 228.4 F 0.0
20:00 229.7 F 0.0
25:00 229.3 F 0.0
30:00 228.4 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 19:47
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:36
CYCLE COMPLETE

03/06/2023 12:47 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____