

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2279

Name: RITTER 1  
Sterilizer ID: V2178198  
09/02/2022 07:26 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.7 F 0.0  
2:00 76.3 F 0.0  
4:00 85.3 F 0.0  
6:00 130.3 F 0.2  
8:00 144.7 F 1.5  
10:00 167.9 F 4.7  
12:00 187.7 F 8.3  
14:00 203.8 F 11.9  
16:00 217.1 F 4.3  
18:00 224.4 F 8.2  
20:00 234.9 F 12.2  
22:00 235.2 F 8.5  
24:00 243.4 F 11.8  
26:00 244.1 F 12.1  
28:00 255.5 F 18.1  
30:00 265.7 F 24.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.1  
1:00 271.8 F 28.6  
2:00 272.1 F 28.5  
3:00 271.4 F 28.2  
4:00 272.4 F 28.9  
  
Min 271.2 F 28.0  
Max 272.5 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.8 F 0.0  
5:00 215.9 F 0.0  
10:00 220.7 F 0.0  
15:00 228.6 F 0.0  
20:00 229.5 F 0.0  
25:00 228.2 F 0.0  
30:00 226.0 F 0.0  
Drying Complete

FILLING: 0:45  
HEATING: 31:43  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 01:09:40  
CYCLE COMPLETE

09/02/2022 08:35 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_