
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001153
12:45 06/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:18
PRESSURIZING 4:19
STERILIZING 5:34
136.0 C 220kPa 5:34
136.3 C 223kPa 6:05
136.2 C 221kPa 6:35
136.2 C 223kPa 7:05
136.3 C 224kPa 7:35
136.2 C 222kPa 8:05
136.4 C 224kPa 8:35
136.2 C 222kPa 9:05
136.2 C 222kPa 9:05
VENTING 9:05
AIR DRYING 10:00
CYCLE COMPLETE 16:08

Digital Signature #
4DC6972E7A62CE49
