

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4498

Name: RITTER2  
Sterilizer ID: V2313228  
04/09/2025 09:43 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.2 F 0.0  
2:00 85.0 F 0.0  
4:00 84.3 F 0.6  
6:00 110.0 F 2.6  
8:00 141.9 F 5.5  
10:00 169.0 F 8.9  
12:00 189.3 F 12.5  
14:00 206.9 F 3.4  
16:00 221.4 F 6.7  
18:00 232.5 F 10.3  
20:00 231.7 F 9.1  
22:00 238.1 F 10.5  
24:00 236.8 F 8.6  
26:00 247.2 F 14.2  
28:00 258.0 F 20.2  
30:00 268.1 F 26.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.0  
1:00 273.0 F 29.6  
2:00 271.6 F 29.0  
3:00 272.8 F 29.6  
4:00 271.6 F 28.9

Min 271.5 F 28.8  
Max 273.0 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.5 F 0.0  
5:00 198.7 F 0.0  
10:00 189.5 F 0.0  
15:00 187.8 F 0.0  
20:00 184.4 F 0.0  
25:00 183.0 F 0.0  
30:00 183.4 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 31:20  
STERILIZING: 4:00  
VENTING: 3:36  
DRYING: 30:00  
TOTAL CYCLE: 01:09:47  
CYCLE COMPLETE

04/09/2025 10:52 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_